



### **Model Number 3 Ply Respirator**

Applicable Standard: KN95 China GB2626-2006 and FFP2 Europe EN 149-2001

Filter Performance:  $\geq 95\%$

Test Agent: NaCl and paraffin oil

Flow Rate: 95 L/min

TIL:  $\leq 8\%$  leakage

Inhalation Resistance:  $\leq 70$  Pa @ 30 L/min

$\leq 240$  Pa @ 95 L/min

$\leq 500$  Pa @ clogging

Exhalation Resistance:  $\leq 300$  Pa

CO2 Clearance Requirement:  $\leq 1\%$

